

培正中學添置教學設備標書1

編號	描述	數量	單價	金額
1	Brand: BIO-RAD The transilluminator is intended for clinical and research laboratory use with DNA stained for UV activation. Wavelength: 365nm Compatible gel type: mini gels Compatible gel size: 50x60mm, 110x60mm, 125x60mm Operation temperature: up to 40 deg C Operation humidity: up to 80%	1		
2	Brand: BIO-RAD Benchtop Shaking Incubator for teaching laboratories doing liquid culture work Up to 4 L media capacity in a small footprint Incubator chamber temperatures: up to 70°C Environmental temperatures: up to 65°C Digital readout of all parameters — temperature, shaker speed, and time Chamber temperature precisely maintained to within 0.5°C of set temperature Shaking platform is compatible with various flask clamps and test tube racks Allows simultaneous incubation of stationary culture plates and shaking liquid cultures	1		
			合計	

報價注意事項

- 1 請用澳門幣為貨幣單位。
- 2 有效期必須超越2020年6月30日並請列明交貨期限。
- 3 請按本校提供之編號及順序報價,未能提供報價之項目請留空。
- 4 請註明是否可作部份項目的購買。
- 5 可提供同類型產品。
- 6 如有需要校方有權要求提供貨板。
- 7 報價單請於2020年5月27日下午5時前，密封後交郭敬文副校長收，信封面請列明投標項目。